

DECLARATION OF CONFORMITY

Hereby :

Société TRIONYX s.a.s.
138, avenue Henri BARBUSSE
92140 CLAMART (France)

Certify :

Safety cabinet – Range 7.90+

Reference :

791+E – 792+E – 793+E – 794+E – 795+E – 798+E
795+900E

Are **tested and homologated 90 minutes according to the standard EN 14470-1 :2004** and to **FM 6050** (Factory Mutual).

These products have a **fire resistance of 105 minutes.**


Standardized labels are in accordance with **ISO 3864, ISO 7010, 92/58/CEE European Directive** and **EN 14470-1.**

Moreover, the manufacture quality is controlled before expedition according to **ISO 9001** process and procedure.

To whom it may concern.

Clamart, the 30th of March 2021


Jean-Philippe THIEULIN
General Manager


Cendrine SERRE-DARGNAT
Quality Manager