

DECLARATION OF CONFORMITY

Hereby : **Société TRIONYX s.a.s.**
138, avenue Henri BARBUSSE
92140 CLAMART (France)

Certify that : **EN 14470-1 Safety cabinets for flammables – Range 7030 and Range 3030**

References :
3031 – 3031E

are tested and homologated 30 minutes to the **EN 14470-1 (04/2004)** and **FM 6050** standards.

Standardized labels are in accordance with **ISO 3864** and **NF X08.003** recommandations, **92/58/CEE** European Directive and **EN 14470-1 (04/2004)** regulation.

Moreover, the manufacture quality is controled before expedition according to **ISO 9001-2008 (certificat n° QUAL/2000/15378b)**.

To whom it may concern.

CLAMART, June 22th, 2015

Jean Philippe THIEULIN
General Manager

Sébastien TOUSSAINT
Quality Manager

TRIONYX s.a.s

SIREN 712 050 210 00020

138, AVENUE HENRI BARBUSSE
92140 CLAMART (France)

Tél. : 01 46 45 80 00 - Fax : 01 46 45 85 98

E-mail : trionyx@trionyx.fr